

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 23 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130579 ✓
Name	David Noll	ID#	6520
Date	7-23-09 ✓	Phone #	786-8928
Agency	ANC1	Instrument Location	Anchorage Jail ✓
Alco S/N	X124669 ✓	Target Value	.081 ✓
High Pressure	200	Alco Test Values	
1 st Alco	.082 ✓	2 nd Alco	.082 ✓
Signature	[Signature] ✓		
(OVER)			

[Signature]
7/30/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 24 day of Aug, 2009.

[Signature]
Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130579✓

(CONTINUED FROM FRONT PAGE)

[Signature]

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JULY 23, 2009 ✓

CD 7/30/09

OPERATOR'S NAME:

NOLL

OPERATOR'S NUMBER: 6520

SUBJECT'S LAST NAME:

CALIBRATION

SUBJECT'S FIRST NAME/MI :

VERIFICATION/O

O.L. #: 012345

DEPT/AGENCY: ANCL

CASE/REPORT: 09-99999

TEST TYPE: V

ALCO TARGET VALUE: .081 ✓

ALCO S/N: X124669 ✓

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.78 in

ALCO TARGET .080

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 29.78 in .082 ✓

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 29.78 in .082 ✓

BLANK TEST .000

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

JULY 23, 2009

TIME 01:55

DN
30210
7-23-09

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 42c

BAROMETER: 29.78 in

FLOW DETECTOR: OKAY

02:01 PUMP

02:02 HIGH SPEED: OKAY

02:02 DETECTOR: OKAY

02:02 DETECTOR: OKAY

02:03 FILTERS: OKAY

02:04 FILTERS: OKAY

02:05 QUARTZ STANDARD: OKAY

02:05 QUARTZ STANDARD: OKAY

02:06 CALIBRATION: OKAY